# Megger.

#### **DCM310 Clampmeter** User Guide





#### ▲ Safety Information

To ensure safe operation and service of the Meter, follow these instructions. Failure to observe warnings can result in severe injury or death.

- Avoid working alone so assistance can be rendered.
- To enhance safety, test leads should be disconnected from instrument when not in use.
- Do not use test leads or the Clamp Meter if they look damaged.
- Do not use the Meter if the Clamp Meter is not operating properly or if it is wet.
- Use the Clamp Meter only as specified in the Instruction card or the protection provided by the Clamp Meter might be impaired.
- Special precautions are necessary when operating in situations where exposed live parts at dangerous voltages may be encountered. Personal protective equipment (not supplied with this instrument) should be used.

## Symbols as marked on the Tester and Instruction card

A	Risk of electric shock	
∕∆	See instruction card	
	Equipment protected by double or reinforced insulation	
Ħ	Battery	
Ť	Earth	
CE	Conforms to EU directives	
4	Application around and removal from hazardous live conductors is permitted	
X	Do not discard this product or throw away	

### Maintenance

Do not attempt to repair this Clamp Meter. It contains no user-serviceable parts. Repair or serving should only be performed by qualified personal.

## Cleaning

Periodically wipe the case with a dry cloth and detergent do not use abrasives or solvents.

## Power On / Off

M

100A~

-----

ത്പില്ല

1228

204, 2004 17 - 11 - 49

## **ACA 20A**



Data Hold





### MAX Hold

## ACA 200A



## **Battery Replacement**



## ACA 400A

380-

▲ CAT. III.600V with respect to earth for the jaw. Tactile barrier for Hand Guard.
▲ Do not hold the meter in front of the Tactile Barrier.
400A



#### Specifications 1-1 General Specifications

#### LCD display digits :

2000 digit large scale LCD readout. Measuring rate : 2.5 times / sec. Overrange display :

"OL" is displayed for "20A" and "200A" range, shows the real value when bigger than "400A" range. Low battery indicator :

is displayed. Replace the battery

when the indicator is appears in the display.

Power requirement : 9V PP3/6LR61 battery. Battery type and life : ALKALINE 580 hours. Rechargeable batteries are not suitable for use with this instrument.

#### **1-2 Environmental Conditions**

Indoor Use. Calibration : One year calibration cycle. Operating temperature :  $0^{\circ}C \sim 30^{\circ}C (\leq 80\% \text{ RH})$  $30^{\circ}C \sim 40^{\circ}C (\leq 75\% \text{ RH})$  $40^{\circ}C \sim 50^{\circ}C (\leq 45\% \text{ RH})$ Storage temperature : -20 to  $+60^{\circ}C$ , 0 to 80% RH (batteries not fitted).

**Operating altitude :** 2000m (6562 ft) **Overvoltage category :** 

IEC 61010-1 600V CAT.Ⅲ,

CAT. Ⅲ equipment is designed to protect against the transients in the equipment in fixed installations, such as distribution panels, feeders and short branch circuits and lighting systems in large buildings. Conductor Size : 27mm diameter. **Pollution degree :** 2 EMC: EN 61326-1 Shock vibration : Sinusoidal vibration per MIL-T-28800E (5 ~ 55 Hz, 3g maximum). Dimensions ( $W \times H \times D$ ): 56mm x 180mm x 32mm Weight: 250g including battery. Accessories : Battery (installed), Carrying case and User manual.

#### **1-3 Electrical Specifications**

Accuracy is  $\pm$  (% reading + number of digits) at 23°C  $\pm$  5°C < 80%RH. **Temperature coefficient :** 0.2 x (Specified accuracy) / °C,

 $< 18^{\circ}C, > 28^{\circ}C.$ 

#### **AC Current**

Function	Range	Accuracy
	0~19.99A	±(3.0% + 0.05 A)
A <b>∼</b> (50~60Hz)	20.00~199.9A	±(2.0% + 0.5 A)
/	200~400A	±(2.0% + 5 A)

## **Limited Warranty**

This Meter is warranted to the original purchaser against defects in material and workmanship for 1 year from the date of purchase. During this warranty period, manufacturer will, at its option, replace or repair the defective unit, subject to verification of the defect or malfunction. This warranty does not cover disposable batteries, or damage from abuse, neglect, accident, unauthorized repair. alteration. contamination, or abnormal conditions of operation or handling. Any implied warranties arising out of the sale of this product, including but not limited to implied warranties of merchantability and fitness for a particular purpose, are limited to the above. The manufacturer shall not be liable for loss of use of the instrument or other incidental or consequential damages, expenses, or economic loss, or for any claim or claims for such damage, expense or economic loss. Some states or countries laws vary, so the above limitations or exclusions may not apply to you.

#### **Megger Limited**

Archcliffe Road, Dover, CT17 9EN United Kingdom Tel +44 (0) 1304 502100

Valley Forge Corporate Centre 2621 Van Buren Avenue, Norristown PA 19403, USA Tel +1-979-690-7925

Email sales@megger.com Web www.megger.com